

MAGLUMI β_2 -MG (CLIA)



130204001M



100



**Shenzhen New Industries
Biomedical Engineering Co., Ltd**
4/F, Wearnes Tech Bldg,
Science & Industry Park,
Nanshan, Shenzhen, 518057 CHINA
Tel. + 86-755-86028224
Fax. + 86-755-26654850

Lotus Global Co., Ltd
15 Alexandra Road
London
UK
NW8 0DP
Tel. + 44-20-75868010
Fax. + 44-20-79006187



FOR PROFESSIONAL USE ONLY

Store at 2-8 °C



COMPLETELY READ THE INSTRUCTIONS BEFORE
PROCEEDING



SYMBOLS EXPLANATIONS



Authorized Representative in the
European community



Manufacturer



Consult instructions for use



Contents of kit



In vitro diagnostic medical device




Batch code



Catalogue number



Use by



Temperature limitation
(store at 2-8 °C)



Sufficient for



Keep away from sunlight



Keep upright for storage

INTENDED USE

The kit has been designed for the quantitative determination of β_2 -microglobulin (β_2 -MG) in human serum or urine.

The method can be used for samples over the range of 0.03-10 μ g/ml.

The test has to be performed on the Fully-auto chemiluminescence immunoassay (CLIA) analyzer MAGLUMI (Including Maglumi 600, Maglumi 1000, Maglumi 1000 Plus, Maglumi 2000, Maglumi 2000 Plus, Maglumi 3000 and Maglumi 4000).

SUMMARY AND EXPLANATION OF THE TEST

β_2 -MG was first isolated in 1968 from the urine of patients with Wilson's disease and cadmium poisoning. It has since been identified as the light chains of the HLA-A, -B, and -C major histocompatibility complex antigens, 100 amino acids in length and noncovalently associated with the heavy chain. In structure and amino acid sequence, it resembles the CH₃ region of IgG, though it is antigenically distinct. β_2 -MG occurs on the surface of nucleated cells - abundantly on lymphocytes and monocytes and on many tumor cell lines. Its function is unknown, but it may control the expression and biosynthesis of antigens on the cell surface.

Because of its low molecular weight (11,800 daltons), 95 percent of all free β_2 -MG is rapidly eliminated by glomerular filtration. Proximal tubular cells then take up 99.9 percent of this filtered amount by endocytosis, after which degradation to amino acids occurs. Normal urinary excretion of β_2 -MG is less than 370 micrograms per 24 hours; higher rates are interpreted as evidence of tubular dysfunction. Increased urinary excretion of β_2 -MG has been observed in a wide variety of conditions including Wilson's disease, Fanconi's syndrome, untreated congenital galactosemia, nephrocalcinosis, cystinosis, chronic potassium depletion, interstitial nephritis, connective-tissue diseases such as rheumatoid arthritis and Sjogren's syndrome, occupational exposure to heavy metals such as cadmium and mercury, upper urinary tract infections, kidney transplantation, and nephrotoxicity resulting from cyclosporine, aminoglycoside or cisplatin therapy.

Elevated serum concentrations in the presence of a normal glomerular filtration rate suggest increased β_2 -MG production or release. Increased levels may be seen in lymphoproliferative diseases such as multiple myeloma, cell chronic lymphocytic leukemia, Hodgkin's disease, non-Hodgkin's lymphoma; systemic lupus erythematosus; rheumatoid arthritis; Sjogren's syndrome; Crohn's disease; and certain viral infections, including cytomegalovirus, non-A and non-B hepatitis and infectious mononucleosis. Elevated serum levels have also been observed in some hemodialysis patients and in renal transplant rejection.

Measurement of β_2 -MG is considered a sensitive means for diagnosing proximal tubular dysfunction. It is reportedly the most reliable test for distinguishing upper from lower urinary tract infections, and a useful method for assessing the results of therapy and diagnosing recurrences of acute pyelonephritis using serial determinations.

PRINCIPLE OF THE TEST

Competitive immunoluminometric assay;

Use an anti- β_2 -MG monoclonal antibody to label ABEI, and use purified β_2 -MG antigen to label FITC. Sample, Calibrator or Control with ABEI Label, FITC Label and magnetic microbeads coated with anti-FITC are mixed thoroughly and incubated at 37°C, forming complexes; after sediment in a magnetic field, decant the supernatant, then cycle washing it for 1 time. Subsequently, the starter reagents are added and a flash chemiluminescent reaction is initiated. The light signal is measured by a photomultiplier as

RLU within 3 seconds and is proportional to the concentration of β_2 -MG present in samples.

CONT

KIT COMPONENTS

Material Supplies

Reagent Integral for 100 determinations	
Nano magnetic microbeads: TRIS buffer, 1.2 % (W/V), 0.2%NaN ₃ , coated with sheep anti-FITC polyclonal antibody.	2.5ml
Calibrator Low: bovine serum, 0.2%NaN ₃	2.5ml
Calibrator High: bovine serum, 0.2%NaN ₃	2.5ml
FITC Label: purified β_2 -MG antigen labeled FITC contains BSA, 0.2%NaN ₃ .	10.5ml
ABEI Label: anti- β_2 -MG monoclonal antibody labeled ABEI, contains BSA, 0.2%NaN ₃ .	10.5ml
Diluent: 0.9%NaCl.	25ml

All reagents are provided ready-to-use.

Reagent Vials in kit box	
Internal Quality Control: containing BSA, 0.2%NaN ₃ . (target value refer to Quality Control Information date sheet)	2.0ml

Internal quality control is only applicable with MAGLUMI system. Instructions for use and target value refer to Quality Control Information date sheet. User needs to judge results with their own standards and knowledge.

Accessories Required But Not Provided

MAGLUMI Reaction Module	REF: 630003
MAGLUMI Starter 1+2	REF: 130299004M
MAGLUMI Wash Concentrate	REF: 130299005M
MAGLUMI Light Check	REF: 130299006M

Please order accessories from SNIBE or our representative.



Preparation of the Reagent Integral

Before the sealing is removed, gentle and careful horizontal shaking of the Reagent Integral is essential (avoid foam formation!) Remove the sealing and turn the small wheel of the magnetic microbeads compartment to and fro, until the colour of the suspension has changed into brown. Place the Integral into the reagent area and let it stand there for 30 min. During this time, the magnetic microbeads are automatically agitated and completely resuspended.

Do not interchange integral component from different reagents or lot!

Storage and Stability

- Sealed: Stored at 2-8°C until the expiry date.
- Opened: Stable for 4 weeks. To ensure the best kit performance, it is recommended to place opened kits in the refrigerator if it's not going to be used on board during the next 12 hours.



- Keep upright for storage.



- Keep away from sunlight.

CALIBRATION AND TRACEABILITY

1) Traceability

To perform an accurate calibration, we have provided the test calibrators standardized against the SNIBE internal reference substance.

2) 2-Point Recalibration

Via the measurement of calibrators, the predefined master curve is adjusted (recalibrated) to a new, instrument-specific measurement level with each calibration.

3) Frequency of Recalibration

- After each exchange of lot (Reagent Integral or Starter Reagents).
- Every week and/or each time a new Integral is used (recommendation).
- After each servicing of the Fully-auto chemiluminescence immunoassay (CLIA) analyzer MAGLUMI..
- If controls are beyond the expected range.
- The room temperature has changed more than 5 °C (recommendation)

SPECIMEN COLLECTION AND PREPARATION

Sample material: serum or urine

Collect samples using standard procedures.

Store at 2-8°C: 12 hours; for longer storage periods: freeze to below -20°C.

Avoid repeated freezing and thawing cycles, stored samples should be thoroughly mixed prior to use (Vortex mixer).

Please ask local representative of SNIBE for more details if you have any doubt.

Vacuum Tubes

- Blank tubes are recommended type for collecting samples.
- Please ask SNIBE for advice if special additive must be used in sample collecting.

Specimen Conditions

- Use caution when handling patient specimens to prevent cross contamination. Use of disposable pipettes or pipette tips is recommended.
- Inspect all samples for bubbles. Remove bubbles with an applicator stick prior to analysis. Use a new applicator stick for each sample to prevent cross contamination.

Preparation for Analysis

- Specimens must be mixed **thoroughly** after thawing by **low** speed vortexing or by gently inverting. Multiple freeze-thaw cycles of specimens should be avoided.
- All samples (patient specimens or controls) should be tested within 3 hours of being placed on board the MAGLUMI System. Refer to the SNIBE service for a more detailed discussion of onboard sample storage constraints.

Storage

- If testing will be delayed for more than 8 hours, remove serum from the serum separator, red blood cells or clot. Specimens removed from the separator gel, cells or clot may be stored up to 12 hours at 2-8°C.
- Specimens can be stored up to 30 days frozen at -20°C or colder.

Shipping

Before shipping serum specimens, it is recommended that specimens be removed from the serum separator, red blood cells or clot. When shipped, specimens must be packaged and labeled in compliance with applicable state, federal and international regulations covering the transport of clinical specimens and infectious substances. Specimens must be shipped frozen (dry ice). Do not exceed the storage time limitations identified in this section of the package insert.

WARNING AND PRECAUTIONS FOR USERS



- For use in *IN-VITRO* diagnostic procedures only.
- Package insert instructions must be carefully followed. Reliability of assay results cannot be guaranteed if there are any deviations from the instructions in this package insert.

Safety Precautions

CAUTION: This product requires the handling of human specimens.

- The calibrators in this kit are prepared from bovine serum products. However, because no test method can offer complete assurance that HIV, Hepatitis B Virus or other infectious agents are absent; these reagents should be considered a potential biohazard and handled with the same precautions as applied to any serum or plasma specimen.
- All samples, biological reagents and materials used in the assay must be considered potentially infectious. They should therefore be disposed of in accordance with the prevailing regulations and guidelines of the agencies holding jurisdiction over the laboratory, and the regulations of each country. Disposable materials must be incinerated; liquid waste must be decontaminated with sodium hypochlorite at a final concentration of 5% for at least half an hour. Any materials to be reused must be autoclaved using an overkill approach. A minimum of one hour at 121°C is usually considered adequate, though the users must check the effectiveness of their decontamination cycle by initially validating it and routinely using biological indicators.
- It is recommended that all human sourced materials be considered potentially infectious and handled in accordance with the OSHA Standard on Blood borne Pathogens¹³. Biosafety Level 214 or other appropriate biosafety practices should be used for materials that contain or are suspected of containing infectious agents.
- This product contains Sodium Azide; this material and its container must be disposed of in a safe way.
- Safety data sheets are available on request.

Handling Precautions

- Do not use reagent kits beyond the expiration date.
- Do not mix reagents from different reagent kits.
- Prior to loading the Reagent Kit on the system for the first time, the microbeads requires mixing to re-suspend microbeads that have settled during shipment.
- For microbeads mixing instructions, refer to the KIT COMPONENTS, Preparation of the Reagent Integral section of this package insert.
- To avoid contamination, wear clean gloves when operating with a reagent kit and sample.
- Over time, residual liquids may dry on the kit surface, please pay attention the silicon film still exists on the surface of the kit.
- For a detailed discussion of handling precautions during system operation, refer to the SNIBE service information.

TEST PROCEDURE

To ensure proper test performance, strictly adhere to the operating instructions of the Fully-auto chemiluminescence immunoassay (CLIA) analyzer MAGLUMI. Each test parameter is identified via a RFID tag on the Reagent Integral. For further information please refer to the Fully-auto chemiluminescence immunoassay (CLIA) analyzer MAGLUMI Operating Instructions.

10µl	Sample, calibrator
+80µl	ABEI Label
+80µl	FITC Label
+20µl	Nano magnetic microbeads
15 min	Incubation
400µl	Cycle washing

DILUTION

Samples with concentrations above the measuring range can be diluted. After manual dilution, multiply the result by the dilution factor. After dilution by the analyzers, the analyzer software automatically takes the dilution into account when calculating the sample concentration.

Availability of sample dilution by analyzer please refers to the MAGLUMI analyzer user software program. Dilution settings please follow MALGUMI analyzer operating instructions.

QUALITY CONTROL

- Observe quality control guidelines for medical laboratories
- Use suitable controls for in-house quality control. Controls should be run at least once every 24 hours when the test is in use, once per reagent kit and after every calibration. The control intervals should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined ranges. Each laboratory should establish guidelines for corrective measures to be taken if values fall outside the range.

LIMITATIONS OF THE PROCEDURE

1) Limitations

A skillful technique and strict adherence to the instructions are necessary to obtain reliable results. Bacterial contamination of samples or repeated freeze-thaw cycles may affect the test results. Assay results should be utilized in conjunction with other clinical and laboratory data to assist the clinician in making individual patient management decisions.

2) Interfering Substances

No interference with test results is seen by concentrations of bilirubin<0.125mg/ml, haemoglobin<500mg/dl or triglycerides<12.5mg/ml.

3) HAMA

Patient samples containing human anti-mouse antibodies (HAMA) may give falsely elevated or decreased values. Although HAMA-neutralizing agents are added, extremely high HAMA serum concentrations may occasionally influence results.

RESULTS

1) Calculation of Results

The analyzer automatically calculates the β₂-MG concentration in each sample by means of a calibration curve which is generated by a 2-point calibration master curve procedure. The results are expressed in µg/ml. For further information please refer to the Fully-auto chemiluminescence immunoassay (CLIA) analyzer MAGLUMI Operating Instructions.

2) Interpretation of Results

- Results of study in clinical centers with group of individuals, 95% of the results were
 - Serum: 0.9-2.7 µg/ml
 - random urine: <0.195 µg/ml
- Results may differ between laboratories due to variations in population and test method. If necessary, each laboratory should establish its own reference range.

PERFORMANCE CHARACTERISTICS

1) Precision

Intra-assay coefficient of variation was evaluated on 2 different levels of control serum repeatedly measured 20 times in the same run, calculating the coefficient of variation.

Intra-assay precision			
Control	Mean(µg/ml)	SD(µg/ml)	CV%

Level 1	1.09	0.05	5.86
Level 2	5.12	0.58	5.29

Inter-assay coefficient of variation was evaluated on three batches of kits. Repeatedly measured 2 different levels of control serum 21 times, calculating the coefficient of variation.

Inter-assay precision			
Control	Mean($\mu\text{g/ml}$)	SD($\mu\text{g/ml}$)	CV%
Level 1	1.08	0.09	9.83
Level 2	5.03	0.48	9.65

2) Analytical Sensitivity

The sensitivity is defined as the concentration of β_2 -MG equivalent to the mean RLU of 20 replicates of the zero standard plus two standard deviations corresponding to the concentration from the standard curve. The sensitivity is typically less than 0.03 $\mu\text{g/ml}$.

3) Specificity

The specificity of the β_2 -MG assay system was assessed by measuring the apparent response of the assay to various potentially cross reactive analytes.

Compound	Concentration	Cross reactivity
IgG	80 $\mu\text{g/ml}$	1%
IgA	40 $\mu\text{g/ml}$	1%
IgM	40 $\mu\text{g/ml}$	1%
IgE	320IU/ml	0.94%

4) Recovery

Consider calibrator high of known concentration as a sample, dilute it by 1:2 ratios with diluents, and measure its diluted concentration for 10 times. Then calculate the recovery of measured concentration and expected concentration. The recovery should be within 90% -110%.

Expected	Mean Measuring	Recovery
2.758 $\mu\text{g/ml}$	2.594 $\mu\text{g/ml}$	94%

5) Linearity

Use β_2 -MG calibrator to prepare the six-point standard curve, measuring all points' RLU except point A, and then do four-parameter linear fitting in double logarithm coordinate, the absolute linear correlation coefficient(r) should be bigger than 0.9800.

Calibrator Point	Concentration $\mu\text{g/ml}$	Absolute linear correlation coefficient (r)
A	0	
B	0.1	r=0.9915
C	0.25	
D	0.5	
E	1	
F	10	

6) Method comparison

A comparison of MAGLUMI β_2 -MG (y) with a commercially available β_2 -MG test (x) using clinical samples gave the following correlations ($\mu\text{g/ml}$):

Linear regression
 $y = 1.1022x + 0.0648$
 $r = 0.9852$

Number of samples measured: 100
The sample concentrations were between 0.67 and 33.53 $\mu\text{g/ml}$.

REFERENCES

- Bethea M, Forman DT. β_2 -Microglobulin: Its Significance and Clinical Usefulness. *Ann Clin Lab Sci* 1990;20(3): 163-168.
- Hallek M et al. Serum β_2 -Microglobulin and Serum Thymidine Kinase are Independent Predictors of Progression-Free, survival in Chronic Lymphocytic Leukemia and

- Immunocytoma. *Leuk Lymphoma* 1996; 22:439-447.
- Johnson PWM et al. β_2 -Microglobulin: A prognostic factor in diffuse aggressive non Hodgkin's lymphomas. *Br J Cancer* 1993; 67:792-797.
- Winchester JF, Salsberg JA, Levin NW (2004). "Beta-2 microglobulin in ESRD: an in-depth review". *Advances in renal replacement therapy* 10 (4): 279-309.
- Gattoni-Celli S, Kirsch K, Timpane R, Isselbacher KJ (1992). "Beta 2-microglobulin gene is mutated in a human colon cancer cell line (HCT) deficient in the expression of HLA class I antigens on the cell surface". *Cancer Res.* 52 (5): 1201-4
- Güssow D, Rein R, Ginjaar I, Hochstenbach F, Seemann G, Kottman A, Ploegh HL (1 November 1987). "The human beta 2-microglobulin gene. Primary structure and definition of the transcriptional unit". *J. Immunol.* 139 (9): 3132-8
- Cunningham BA, Wang JL, Berggård I, Peterson PA (November 1973). "The complete amino acid sequence of beta 2-microglobulin". *Biochemistry* 12(24): 4811-22.
- Hönig R, Marsen T, Schad S et al. (1989). "Correlation of beta-2-microglobulin concentration changes to changes of distribution volume". *The International journal of artificial organs* 11 (6): 459-64.